

EC DECLARATION OF CONFORMITY

Laprosurge Ltd.
73a High Road,
Bushey Heath,
Hertfordshire WD23 1EL.

Declares that the following medical device:

EC10SLB Auto 10mm BLADELESS TROCAR and CANNULA
with gas tap and automatic reducing valve.

has been classified as Class 2a and is in conformity with the essential requirements of Council Directive 93/42/EEC and is subject to the procedure set out in Annex 2 of the Directive under the Supervision of SGS United Kingdom Ltd. being the Notified Body number 0120.



Signed: _____ Date: 4 February 2011

David Campbell
Director Laprosurge Ltd.